

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00506	Calib. Date:	12/03/2019
Certification File No.:	00507	Cert. Date:	12/03/2019
Linearity File No.:	00508	Lin. Date:	12/03/2019
Solution File No.:	00525	Soln. Date:	03/22/2020
Sequential File No.:	00525	File Date:	03/22/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	19110	Expires:	04/09/2021
		Bottle No.:	0650

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:55S		
Control 1 EC	0.098%	09:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	09:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56S		
Control 2 EC	0.097%	09:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:57S		
Control 3 EC	0.097%	09:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	09:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:59S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: 

Badge No.: 115

Date: 03/22/2020