

# Calibrating Unit

## New Standard Solution Report

|                         |                       |              |              |
|-------------------------|-----------------------|--------------|--------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.:  | ARXB-0076    |
| Location:               | SPRING LAKE POLICE    |              |              |
| Calibration File No.:   | 00530                 | Calib. Date: | 05/14/2020   |
| Certification File No.: | 00531                 | Cert. Date:  | 05/14/2020   |
| Linearity File No.:     | 00532                 | Lin. Date:   | 05/14/2020   |
| Solution File No.:      | 00539                 | Soln. Date:  | 08/02/2020   |
| Sequential File No.:    | 00539                 | File Date:   | 08/02/2020   |
| Calibrating Unit:       | WET                   | Model No.:   | CU-34        |
| Control Solution %:     | 0.100%                | Serial No.:  | DDYB S3-0002 |
| Solution Control Lot:   | 19150                 | Expires:     | 06/10/2021   |
|                         |                       | Bottle No.:  | 0430         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 22:15D |                |                     |
| Control 1 EC      | 0.097% | 22:16D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.097% | 22:16D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:16D |                |                     |
| Control 2 EC      | 0.096% | 22:17D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.098% | 22:17D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:18D |                |                     |
| Control 3 EC      | 0.096% | 22:18D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.097% | 22:18D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:19D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: Pat P. Gilroy 115

Badge No.: 115

Date: 08/02/2020