

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00530	Calib. Date:	05/14/2020
Certification File No.:	00531	Cert. Date:	05/14/2020
Linearity File No.:	00532	Lin. Date:	05/14/2020
Solution File No.:	00540	Soln. Date:	08/30/2020
Sequential File No.:	00540	File Date:	08/30/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0002
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0301

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:15D		
Control 1 EC	0.097%	22:16D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:16D		
Control 2 EC	0.096%	22:17D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	22:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:18D		
Control 3 EC	0.096%	22:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____

Sgt P. Gilroy 115

Badge No.: 115

Date: 08/30/2020