

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00542	Calib. Date:	09/30/2020
Certification File No.:	00543	Cert. Date:	09/30/2020
Linearity File No.:	00544	Lin. Date:	09/30/2020
Solution File No.:	00549	Soln. Date:	10/12/2020
Sequential File No.:	00549	File Date:	10/12/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0002
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0242

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:49D		
Control 1 EC	0.100%	22:50D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:51D		
Control 2 EC	0.099%	22:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:52D		
Control 3 EC	0.099%	22:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____



Badge No.: 115

Date: 10/12/2020