

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00542	Calib. Date:	09/30/2020
Certification File No.:	00543	Cert. Date:	09/30/2020
Linearity File No.:	00544	Lin. Date:	09/30/2020
Solution File No.:	00556	Soln. Date:	01/16/2021
Sequential File No.:	00556	File Date:	01/16/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0002
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0238

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:57S		
Control 1 EC	0.100%	08:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:58S		
Control 2 EC	0.099%	08:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59S		
Control 3 EC	0.099%	09:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HEINE

First Name: JAMES

MI: E

Signature: _____

PH1 James Hein #121

Badge No.: 121

Date: 01/16/2021