

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00542	Calib. Date:	09/30/2020
Certification File No.:	00543	Cert. Date:	09/30/2020
Linearity File No.:	00544	Lin. Date:	09/30/2020
Solution File No.:	00557	Soln. Date:	02/07/2021
Sequential File No.:	00557	File Date:	02/07/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0002
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0544

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:54S		
Control 1 EC	0.101%	10:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56S		
Control 2 EC	0.099%	10:56S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57S		
Control 3 EC	0.099%	10:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: 

Badge No.: 115

Date: 02/07/2021