

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Spring Lake Borough Police Department 311 Washington Avenue Spring Lake, NJ 07762		2. Contact: Sgt. Patrick Gilroy	4. Date: 7/21/21
		3. Phone Number: 732-449-1234	5. County: Monmouth
6. Alcotest Instrument Serial Number: ARXB-0076	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:

The alcotest instrument was returned from outside evaluation and placed back in service.

10. Comments:

See Draeger Return and Repair form.

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE

Last Known Sequential File #: NA

☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

☐ 12. The above Instrument/Component is placed out of service pending further evaluation.

☒ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dennis J Lutz #7045
Name & Badge Number (Print)

Tpr I Lutz 7045 7-21-21
Signature Date

RETURN AND REPAIR FORM

Dräger

Customer Information

B: _____

S: _____

Company Name: Borough Of Spring Lake, Spring Lake, NJ

Date Received: 06/11/2021

Date given to service: 06/11/2021

Carrier: ☐ FedEx ☐ UPS ☐ USPS

Shipping Method:

☐ GRD

☐ NDA-PRI

☐ 3DAY

☐ NDA-STD

☐ 2DAY

Product: ☐ 6510

☐ 6810

☒ 7110

☐ 8610

☐ 7510

☐ DT5000

☐ 7410 Upper-half

☐ 9510

☐ 7410 Whole

Serial Number: AR XB-0076

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter

☐ Regulator

☐ Printer Ribbon

☐ Printer Paper

☐ Mouthpieces

☐ 9510 Stylus

☐ 9510 Top Cover

☐ Carrying Case

☐ Dry Gas

☐ Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	2.0	
MP Labor	Labor	4.0	
8315075	Main PC Board	1.0	

Repair Notes: Replaced defective main PC board.

CAL W/QC AND OPS CHECK

Service Technician

GR

Date:

6/21/21

Service note

Customer no. 150063898
Number of report 306153510
Date of receiving 06/08/2021
Please reference on inquiries

Customer
BOROUGH OF SPRING LAKE
ATTN: FINANCE DEPARTMENT
423 WARREN AVE
SPRING LAKE NJ 07762-1232

Consignee
BOROUGH OF SPRING LAKE POLICE DEPT
ATTN: CHIEF ROBERT DAWSON
311 WASHINGTON AVE
SPRING LAKE NJ 07762-1431

Your order

Date of order:

Your reference:
ARXB-0076

NL-Text

Maintenance device list

Your contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 144971914
		REPAIR ARXB-0076
		Eq.No. : 1034472370 SN: ARXB-0076
		Material: 8314246 ALCOTEST 7110 MK III USA
		Inv.No. :
		Location:
		REPAIR ARXB-0076
		R001 Repair time - normal hours
		8315075 Best. LP MYC USA, A7110 MKIII
4.0	H	
1	EA	

Delivery receipt



Customer no.
150063898

Number of report
306153510
Please reference on inquiries

Date of receiving Delivery receipt date
/ / 06/30/2021

Consignee, 150063899
BOROUGH OF SPRING LAKE POLICE DEPT
ATTN: CHIEF ROBERT DAWSON
PO: ARXB-0076
311 WASHINGTON AVE
SPRING LAKE NJ 07762-1431

Customer, 150063898
BOROUGH OF SPRING LAKE
ATTN: FINANCE DEPARTMENT
PO: ARXB-0076
423 WARREN AVE
SPRING LAKE NJ 07762-1232

Your order

Date of order:

Purch.ord.no.:
ARXB-0076

Branch text

Draeger contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
001	ALCOTEST 7110 MK III USA Repair free of charge	8314246	1034472370	ARXB-0076

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road, Telford, PA 18969
An Equal Opportunity Employer M / F / V / H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS Lockbox Processing
Lockbox #13369
100 Grove Road
Suite E
West Deptford, NJ 08066

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Spring Lake Borough Police Department 311 Washington Avenue Spring Lake, NJ 07762		2. Contact: Sgt. Patrick Gilroy 3. Phone Number: 732-449-1234	4. Date: 6/2/21 5. County: Monmouth
6. Alcotest Instrument Serial Number: ARXB-0076	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:

I responded to Spring Lake Borough Police Department to place the Alcotest 7110 back in service for the second time. While conducting Part II - Linearity Tests, I received a control test failure due to incorrect control solution %. While preparing for the 3rd control test(the first 0.080% test), the screen read 2.848%. The test subsequently failed. On the printout, the "Control Solution %" reads 2.848%. The alcotest was just returned from Draeger for the same failure.

10. Comments:

Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This is the third time being sent back for the same failure that occurred. The instrument was placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Last Known Sequential File #:00578

☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

☒ 12. The above Instrument/Component is placed out of service pending further evaluation.

☐ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dennis J Lutz #7045

Name & Badge Number (Print)

Tpr I Lutz 7045 6-2-21

Signature

Date

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Borough Of Spring Lake Police Dept., Spring Lake, NJ

Date Received: 04/30/2021

Date given to service: 04/30/2021

Carrier: ☐ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☐ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR XB-0076

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	2.0	
MP Labor	Labor	.5	

Repair Notes: Unit software re-imaged.

CAL W/QC AND OPS CHECK

Service Technician

GR

Date:

5/14/2021

Delivery receipt



Customer no.
150063898

Number of report
306045317
Please reference on inquiries

Date of receiving Delivery receipt date
/ / 05/17/2021

Consignee, 150063899
BOROUGH OF SPRING LAKE POLICE DEPT
ATTN: DSGT. CHRIS KUCIHSKI
311 WASHINGTON AVE
SPRING LAKE NJ 07762-1431

Customer, 150063898
BOROUGH OF SPRING LAKE POLICE DEPT
ATTN: RECORDS
423 WARREN AVE
SPRING LAKE NJ 07762-1232

Your order

Date of order:
04/26/2021
Purch.ord.no.:
ARXB-0076

Branch text

Draeger contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
	ALCOTEST 7110 MK III USA	8314246	1034472370	ARXB-0076
	REPAIR/ ARXB-0076 Repair free of charge			

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road; Telford, PA 18969
An Equal Opportunity Employer M / F / V / H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS Lockbox Processing
Lockbox #13369
100 Grove Road
Suite E
West Deptford, NJ 08066

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

Service note

Customer no.
150063898

Number of report
306045317

Date of receiving

Please reference on inquiries

Customer
BOROUGH OF SPRING LAKE
ATTN: FINANCE DEPARTMENT
423 WARREN AVE
SPRING LAKE NJ 07762-1232

Consignee
BOROUGH OF SPRING LAKE POLICE DEPT
ATTN: CHIEF ROBERT DAWSON
311 WASHINGTON AVE
SPRING LAKE NJ 07762-1431

Your order

Date of order:
04/26/2021
Your reference:
ARXB-0076

NL-Text

Your contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 144734589
		REPAIR/ ARXB-0076
		Eq.No. : 1034472370 SN: ARXB-0076
		Material: 8314246 ALCOTEST 7110 MK III USA
		Inv.No. :
		Location:
		REPAIR/ ARXB-0076
	0.5 H	R001 Repair time - normal hours

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Spring Lake Borough Police Department 311 Washington Avenue Spring Lake, NJ 07762		2. Contact: Sgt. Patrick Gilroy	4. Date: 4/23/21
		3. Phone Number: 732-449-1234	5. County: Monmouth
6. Alcotest Instrument Serial Number: ARXB-0076	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:

I responded to Spring Lake Borough Police Department to place the Alcotest 7110 back in service. While conducting Part II - Linearity Tests, I received a control test failure due to incorrect control solution %. While preparing for the 3rd control test(the first 0.080% test), the screen read 2.848%. The test subsequently failed. On the printout, the "Control Solution %" reads 2.848%. The alcotest was just returned from Draeger for the same failure.

10. Comments:

Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This is the same failure that occurred prior to the repair. The instrument was placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Last Known Sequential File #: 00566

☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

☒ 12. The above Instrument/Component is placed out of service pending further evaluation.

☐ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dennis J Lutz #7045

Name & Badge Number (Print)

Signature

Tpr I Lutz 7045 4-23-21

Date

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Borough Of Spring Lake, Spring Lake, NJ

Date Received: 03/10/2021

Date given to service: 03/10/2021

Carrier: ☐ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☐ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610 Serial Number: AR XB-0076
☐ 7510 ☐ DT5000 Printer Serial#: AR
☐ 7410 Upper-half ☐ 9510 Sim Serial#: DD
☐ 7410 Whole Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	2.0	
MP Labor	Labor	.5	
6808455	Fuel Cell	1	
6808374	Geared Engine	1	

Repair Notes: Replaced defective fuel cell and Geared Engine.

FUEL CELL = ARND-1676

CAL W/QC AND OPS CHECK

Service Technician: GR Date: 04/07/2021

Delivery receipt



Customer no.
150063898

Number of report
305907962
Please reference on inquiries

Date of receiving Delivery receipt date
/ / 04/08/2021

Consignee, 150063898
BOROUGH OF SPRING LAKE POLICE DEPT
ATTN: D SGT CHRIS KUCINSKI
PO#ARXB-0076
311 WASHINGTON AVE
SPRING LAKE NJ 07762-1431

Customer, 150063898
BOROUGH OF SPRING LAKE
ATTN: D SGT CHRIS KUCINSKI
PO#ARXB-0076
423 WARREN AVE
SPRING LAKE NJ 07762-1232

Your order

Date of order:
03/03/2021
Purch.ord.no.:
21-00405

Branch text

Draeger contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
001	ALCOTEST 7110 MK III USA	8314246	1034472370	ARXB-0076

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road; Telford, PA 18969
An Equal Opportunity Employer M / F / V / H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS Lockbox Processing
Lockbox #13369
100 Grove Road
Suite E
West Deptford, NJ 08066

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

Service note

Customer no. 150063898
Number of report 305907962
Date of receiving 03/03/2021
Please reference on inquiries

Customer
BOROUGH OF SPRING LAKE
ATTN: FINANCE DEPARTMENT
423 WARREN AVE
SPRING LAKE NJ 07762-1232

Consignee
BOROUGH OF SPRING LAKE POLICE DEPT
ATTN: CHIEF ROBERT DAWSON
311 WASHINGTON AVE
SPRING LAKE NJ 07762-1431

Your order

Date of order:
03/03/2021
Your reference:
21-00405

NL-Text

Maintenance device list

Your contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 144452613
		REPAIR/ARXB-0076
		Eq.No. : 1034472370 SN: ARXB-0076
		Material: 8314246 ALCOTEST 7110 MK III USA
		Inv.No. :
		Location:
		REPAIR/ARXB-0076
	0.5 H	R001 Repair time - normal hours
	1 EA	6808455 Sensor Alcotest B
	1 EA	6808374 Geared Engine, A7110/A9510

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Spring Lake Borough Police Department 311 Washington Avenue Spring Lake, NJ 07762		2. Contact: Sgt. Patrick Gilroy	4. Date: 03/03/21
		3. Phone Number: 732-449-1234	5. County: Monmouth
6. Alcotest Instrument Serial Number: ARXB-0076	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:

I responded to Spring Lake Borough Police Department to complete the semi-annual recalibration of the Alcotest 7110. While conducting Part II - Linearity Tests, I received a control test failure due to incorrect control solution %. While preparing for the first Linearity Test on the 0.080% solution, the screen read 2.848%. The test subsequently failed.

10. Comments:

Operational condition of the instrument was assessed by reviewing the performance history of the instrument. On a previous AIR (sequential file number 00555) incorrect characters appeared next to Control Solution %. The instrument was placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Last Known Sequential File #: 00561

☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

☒ 12. The above Instrument/Component is placed out of service pending further evaluation.

☐ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dennis J Lutz #7045

Name & Badge Number (Print)

Signature

Tpr I Lutz 7045 3-3-21

Date