

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00600 **Calib. Date:** 07/21/2021 **Calib. No.:** 00041
Certification File No.: 00601 **Cert. Date:** 07/21/2021 **Cert. No.:** 00031
Linearity File No.: 00602 **Lin. Date:** 07/21/2021 **Lin. No.:** 00046
Solution File No.: 00604 **Soln. Date:** 08/25/2021 **Soln. No.:** 00228
Sequential File No.: 00604 **File Date:** 08/25/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDYB S3-0001
Control Solution %: 0.100% **Expires:** 05/04/2023
Solution Control Lot: 21150 **Bottle No.:** 1118

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:01D		
Control 1 EC	0.101%	10:02D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02D		
Control 2 EC	0.100%	10:03D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	10:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04D		
Control 3 EC	0.100%	10:04D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	10:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCINSKI

First Name: CHRISTOPHER

MI: J

Signature: _____

Badge No.: 118

Date: 08/25/2021