

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00600	Calib. Date:	07/21/2021
Certification File No.:	00601	Cert. Date:	07/21/2021
Linearity File No.:	00602	Lin. Date:	07/21/2021
Solution File No.:	00609	Soln. Date:	10/08/2021
Sequential File No.:	00609	File Date:	10/08/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	21150	Expires:	05/04/2023
		Bottle No.:	1056

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:14D		
Control 1 EC	0.103%	01:14D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	01:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:15D		
Control 2 EC	0.102%	01:16D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.103%	01:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:16D		
Control 3 EC	0.102%	01:17D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.103%	01:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PLOSKONKA

First Name: ANTHONY

MI: M

Signature: 

Badge No.: 113

Date: 10/08/2021