

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00600 Calib. Date: 07/21/2021 Calib. No.: 00041
Certification File No.: 00601 Cert. Date: 07/21/2021 Cert. No.: 00031
Linearity File No.: 00602 Lin. Date: 07/21/2021 Lin. No.: 00046
Solution File No.: 00613 Soln. Date: 11/07/2021 Soln. No.: 00231
Sequential File No.: 00613 File Date: 11/07/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0001
Control Solution %: 0.100% Expires: 05/04/2023
Solution Control Lot: 21150 Bottle No.: 1116

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:33S		
Control 1 EC	0.104%	22:34S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.103%	22:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:34S		
Control 2 EC	0.102%	22:35S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.103%	22:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:36S		
Control 3 EC	0.102%	22:36S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.104%	22:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:37S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature:  115

Badge No.: 115

Date: 11/07/2021