

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00619	Calib. Date:	12/28/2021
Certification File No.:	00620	Cert. Date:	12/28/2021
Linearity File No.:	00621	Lin. Date:	12/28/2021
Solution File No.:	00623	Soln. Date:	01/17/2022
Sequential File No.:	00623	File Date:	01/17/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0002
Solution Control Lot:	21150	Expires:	05/04/2023
		Bottle No.:	1058

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:11S		
Control 1 EC	0.100%	22:12S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:12S		
Control 2 EC	0.099%	22:13S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:14S		
Control 3 EC	0.099%	22:14S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: *Pat Gilroy 115*

Badge No.: 115

Date: 01/17/2022