

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00619 Calib. Date: 12/28/2021 Calib. No.: 00042
Certification File No.: 00620 Cert. Date: 12/28/2021 Cert. No.: 00032
Linearity File No.: 00621 Lin. Date: 12/28/2021 Lin. No.: 00047
Solution File No.: 00626 Soln. Date: 02/07/2022 Soln. No.: 00236
Sequential File No.: 00626 File Date: 02/07/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.:]
Control Solution %: 0.100% Expires: 05/04/2023
Solution Control Lot: 21150 Bottle No.: 1003

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:51S		
Control 1 EC	0.098%	16:52S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:52S		
Control 2 EC	0.097%	16:53S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	16:53S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:54S		
Control 3 EC	0.098%	16:54S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: Sgt Pat Gilroy 115

Badge No.: 115

Date: 02/07/2022