

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00619 **Calib. Date:** 12/28/2021 **Calib. No.:** 00042
Certification File No.: 00620 **Cert. Date:** 12/28/2021 **Cert. No.:** 00032
Linearity File No.: 00621 **Lin. Date:** 12/28/2021 **Lin. No.:** 00047
Solution File No.: 00627 **Soln. Date:** 02/21/2022 **Soln. No.:** 00237
Sequential File No.: 00627 **File Date:** 02/21/2022

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:**]
Control Solution %: 0.100% **Expires:** 05/04/2023
Solution Control Lot: 21150 **Bottle No.:** 1140

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:53S		
Control 1 EC	0.099%	10:54S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:54S		
Control 2 EC	0.098%	10:55S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56S		
Control 3 EC	0.099%	10:56S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: Sgt PS Gilroy 115

Badge No.: 115

Date: 02/21/2022