

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00619	Calib. Date:	12/28/2021
Certification File No.:	00620	Cert. Date:	12/28/2021
Linearity File No.:	00621	Lin. Date:	12/28/2021
Solution File No.:	00634	Soln. Date:	04/10/2022
Sequential File No.:	00634	File Date:	04/10/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	]
Solution Control Lot:	21150	Expires:	05/04/2023
		Bottle No.:	1108

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:14D		
Control 1 EC	0.099%	23:15D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:16D		
Control 2 EC	0.098%	23:16D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:17D		
Control 3 EC	0.098%	23:18D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature:

Patrick Gilroy 115

Badge No.: 115

Date: 04/10/2022