

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00636	Calib. Date:	04/21/2022
Certification File No.:	00637	Cert. Date:	04/21/2022
Linearity File No.:	00638	Lin. Date:	04/21/2022
Solution File No.:	00652	Soln. Date:	07/28/2022
Sequential File No.:	00652	File Date:	07/28/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	21150	Expires:	05/24/2023
		Bottle No.:	1014

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:34D		
Control 1 EC	0.099%	13:35D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	13:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:35D		
Control 2 EC	0.098%	13:36D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:37D		
Control 3 EC	0.098%	13:37D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	13:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: PEOSKONKA

First Name: ANTHONY

MI: M

Signature:



Badge No.: 113

Date: 07/28/2022