

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00636	Calib. Date:	04/21/2022
Certification File No.:	00637	Cert. No.:	00043
Linearity File No.:	00638	Cert. Date:	04/21/2022
Solution File No.:	00657	Lin. Date:	04/21/2022
Sequential File No.:	00657	Lin. No.:	00048
		Soln. Date:	08/10/2022
		Soln. No.:	00249
		File Date:	08/10/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	21150	Expires:	05/04/2023
		Bottle No.:	1089

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:11D		
Control 1 EC	0.097%	06:12D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	06:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:12D		
Control 2 EC	0.095%	06:13D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	06:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:13D		
Control 3 EC	0.095%	06:14D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	06:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:15D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PLOSKONKA

First Name: ANTHONY

MI: M

Signature: 

Badge No.: 113

Date: 08/10/2022