

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00636	Calib. Date:	04/21/2022
Certification File No.:	00637	Cert. Date:	04/21/2022
Linearity File No.:	00638	Lin. Date:	04/21/2022
Solution File No.:	00667	Soln. Date:	09/09/2022
Sequential File No.:	00667	File Date:	09/09/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	21150	Expires:	05/04/2023
		Bottle No.:	1049

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:22D		
Control 1 EC	0.098%	18:23D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	18:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:24D		
Control 2 EC	0.096%	18:24D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:25D		
Control 3 EC	0.097%	18:25D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	18:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KELLY	First Name: WILLIAM	MI: E
Signature: 	Badge No.: 124	
	Date: 09/09/2022	