

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00669 Calib. Date: 09/14/2022 Calib. No.: 00044
Certification File No.: 00670 Cert. Date: 09/14/2022 Cert. No.: 00034
Linearity File No.: 00671 Lin. Date: 09/14/2022 Lin. No.: 00049
Solution File No.: 00675 Soln. Date: 10/06/2022 Soln. No.: 00254
Sequential File No.: 00675 File Date: 10/06/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0001
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 0169

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:31D		
Control 1 EC	0.101%	18:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:32D		
Control 2 EC	0.099%	18:33D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:34D		
Control 3 EC	0.099%	18:34D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NOBLE

First Name: WILLIAM

MI: T

Signature: _____

Badge No.: 128

Date: 10/06/2022