

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00669	Calib. Date:	09/14/2022
Certification File No.:	00670	Cert. Date:	09/14/2022
Linearity File No.:	00671	Lin. Date:	09/14/2022
Solution File No.:	00678	Soln. Date:	10/24/2022
Sequential File No.:	00678	File Date:	10/24/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	22070	Expires:	02/17/2024
		Bottle No.:	0139

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:19D		
Control 1 EC	0.101%	02:19D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:20D		
Control 2 EC	0.099%	02:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	02:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:21D		
Control 3 EC	0.099%	02:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PHILLIPS

First Name: MATTHEW

MI: J

Signature:

Matthew Phillips #126

Badge No.: 126

Date: 10/24/2022