

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00636	Calib. Date:	04/21/2022
Certification File No.:	00637	Cert. Date:	04/21/2022
Linearity File No.:	00638	Lin. Date:	04/21/2022
Solution File No.:	00643	Soln. Date:	05/25/2022
Sequential File No.:	00643	File Date:	05/25/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0002
Solution Control Lot:	21150	Expires:	05/04/2023
		Bottle No.:	1022

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:43D		
Control 1 EC	0.100%	22:44D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	22:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:45D		
Control 2 EC	0.098%	22:45D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	22:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:46D		
Control 3 EC	0.098%	22:47D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____

Sgt P. Gilroy 115

Badge No.: 115

Date: 05/25/2022