

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00669	Calib. Date:	09/14/2022
Certification File No.:	00670	Cert. Date:	09/14/2022
Linearity File No.:	00671	Lin. Date:	09/14/2022
Solution File No.:	00683	Soln. Date:	11/28/2022
Sequential File No.:	00683	File Date:	11/28/2022
		Calib. No.:	00044
		Cert. No.:	00034
		Lin. No.:	00049
		Soln. No.:	00257
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	21480	Expires:	12/28/2023
		Bottle No.:	1318

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:57S		
Control 1 EC	0.101%	16:57S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	16:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:58S		
Control 2 EC	0.101%	16:59S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	16:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:59S		
Control 3 EC	0.101%	17:00S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	17:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PHILLIPS

First Name: MATTHEW

MI: J

Signature:

Matthew Phillips #126

Badge No.: 126

Date: 11/28/2022