

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00685	Calib. Date:	12/14/2022
Certification File No.:	00686	Cert. Date:	12/14/2022
Linearity File No.:	00687	Lin. Date:	12/14/2022
Solution File No.:	00693	Soln. Date:	01/17/2023
Sequential File No.:	00693	File Date:	01/17/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	21480	Expires:	12/28/2023
		Bottle No.:	1380

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:53S		
Control 1 EC	0.101%	17:54S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:55S		
Control 2 EC	0.099%	17:55S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:56S		
Control 3 EC	0.099%	17:57S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PHILLIPS

First Name: MATTHEW

MI: J

Signature: 

Badge No.: 126

Date: 01/17/2023