

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00685	Calib. Date:	12/14/2022
Certification File No.:	00686	Cert. No.:	00045
Linearity File No.:	00687	Cert. Date:	12/14/2022
Solution File No.:	00699	Lin. Date:	12/14/2022
Sequential File No.:	00699	Soln. Date:	02/26/2023
		File Date:	02/26/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	21480	Expires:	12/28/2023
		Bottle No.:	1482

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:31S		
Control 1 EC	0.100%	23:32S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:32S		
Control 2 EC	0.100%	23:33S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:34S		
Control 3 EC	0.100%	23:34S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PHILLIPS

First Name: MATTHEW

MI: J

Signature: *Matthew Phillips #126*

Badge No.: 126

Date: 02/26/2023